



ИЗПЪЛНИТЕЛНА АГЕНЦИЯ ПО ОКОЛНА СРЕДА - СОФИЯ

РЕГИОНАЛНА ЛАБОРАТОРИЯ - ПЛОВДИВ

Ул. "Перущица" №1, e-mail: ecolab\_plovdiv@mail.bg, Телефон/Факс: 032/642 723

Sheet 1

Total number of sheets 2

## TESTING PROTOCOL

**No. 1300/2010**

**Date: 08.10.2010**

Accredited testing laboratory

Certificate No 69LI/valid until 30.11.2012

Issued by the Bulgarian Accreditation Service - Executive Agency at the  
Ministry of Economy, Energy and Tourism

### 1. Air Emissions

Pobeda village (41°26,122'N; 25°38,450'E):

3339.10 - from 16 hrs. on 28.09.10 till 15.30 hrs. on 29.09.10;

(name of product; type of component for testing and describing; sampling site)

**2. Company which requested the testing:** BALKAN MINERAL AND MINING EAD, Contract No. 4/27.09.2010, Order No. 14/27.09.2010, Sampling protocol No. 103/2010

(name of the Company, No. and date of the sampling protocol )

**3. Method of testing:** BNS EN 12341, BNS EN 14907

(name and no. of the standards or validated methods)

**4. Date of receiving of samples for testing in the laboratory:** 04/10/2010

**5. Quantity of tested samples:** 1 item /filters/

(quantity of samples/ and their mass)

**6. Date/period of the testing:** 04.10.2010 - 08.10.2010

**MANAGER OF THE INDEPENDENT LABORATORY:**

**Eng. Stanka Chokova**

(family name, signature, seal)

## 7. TEST RESULTS:

#	Type of testing/characteristics	Unit of measurement	Standards/validated methods	No. of sample  acc. to incoming-outgoing logbook	TEST RESULTS: /value, below detection/	Parameter value and tolerance#	Test conditions	Deviations from the method of testing
1	2	3	4	5	6	7	8	9
1	PM <sub>10</sub>	µg/m <sup>3</sup>	BNS EN 12341, BNS EN 14907	3337.10	10±1	50	t° -19,0 ÷21,0 °C Rel. humidity - 45 ÷55 %	No deviations from the testing method

**LEGEND:** # - pursuant to Regulations No. 12/15.07.2010, SG No. 58/2010

**NOTE:** The test results refer only to the tested samples.

Excerpts of the protocol of testing should not be copied without the written consent of the testing laboratory.

### PERSONS WHO CARRIED OUT THE TESTING:

**Eng. D. Karadzhova**  
 (family name, signature)

### MANAGER OF THE INDEPENDENT LABORATORY

**Eng. Stanka Chokova**  
 (family name, signature, seal)